

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME**CB TEST CERTIFICATE**

Product	Aerosol dosimeter for LDSA
Name and address of the applicant	naneos particle solutions GmbH Alte Spinnerei 9, 5210 Windisch Switzerland
Name and address of the manufacturer	naneos particle solutions GmbH Alte Spinnerei 9, 5210 Windisch Switzerland
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	Naneos particle solutions GmbH Alte Spinnerei 9, 5210 Windisch Switzerland
Ratings and principal characteristics	Device: 5 V DC / 1.5 A Supplementary power supply: Input: 100-240V, 50/60Hz, 0.3A; Output: 5V DC, 1.5A, 7.5W
Trademark / Brand (if any)	naneos
Customer Test Facility (CTF) Stage used	
Model / Type Ref.	Partector 2, Part No: 171
Additional information (if necessary may also be reported on page 2)	
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
As shown in the Test Report Ref. No. which forms part of this Certificate	22CH-01062.S01

This CB Test Certificate is issued by the National Certification Body



eurofins

Eurofins Electric & Electronic Product Testing AG
Luppenstrasse 3
8320 Fehraltorf
SWITZERLAND

Date: 2023-05-24

Signature: Clemens Padel 